



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 200 Sample  
Lot Number 24228025

Ref Number 53360  
Exp Date 2026-02-06  
Date of Manufacture 2024-08-06  
UDI# (01)00385480533607(11)240806(17)260206(10)24228025

Cal A	L/N	24219027		Cal B	L/N	24219028	
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
	pH	7.359	N/A		pH	6.837 - 6.843	Pass
pCO2	mmHg	24.4	N/A	pO2	mmHg	167.0	N/A
Flush	L/N	24219024		Reference	L/N	24219029	
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
pCO2	mmHg	51.6	N/A	KCl		1.8 - 2.2	Pass
pO2	mmHg	99.2	N/A				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Haneen Aldarwish*

Approval

QC Inspector  
Title

08/27/2024  
Date