OI-922

Page 39 of 66

Rev. BK

Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:		LIFESCAN	
Description:		TEST STRIPS, STATSTRIP <sup>™</sup> GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT	
Lifescan Purchase Order No.:		642807	
Catalog Number:		42214	
Nova Biomedical P/N		54515	
One Touch P/N		02315302	
Target Range (mg/dL)	SD(mg/dL) /CV (%)		
30-50	≤4 mg/dL	PASS / FAIL	
51-110	≤ 6.0%	PASS FAIL	
111-150	≤ 5.0%	PASS / FAIL	
151-250	≤ 5.0%	PASS FAIL	
251-400	≤ 5.0%	PASS/ FAIL	
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%		PASS / FAIL	
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method ≥ 98%		PASS / FAIL	
Qty:		24,192	
Lot No.		0324206249	
Date of Mfg. (DOM):		2024 / 07 / 24 (YYYY/MM/DD)	
Expiration Date:		2026 / 07 / 25 (YYYY/MM/DD)	
Unique Device Identifier (UDI):		(01)20385480422141 (11)240724 (17)260725	
1		(10) 0 3 2 4 2 0	6249

Quality Control Manager

8/27/24 Date

Nova Biomedical Corporation

200 Prospect St

Waltham, MA 02454