

CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge Lot Number 24235015

tridge Ref Number 54215 Exp Date 2025-08-22 Date of Manufacture 2024-08-22 UDI# (01)00385480542159(11)240822(17)250822(10)24235015

Ctl Level 1 Ctl Level 2

 Spec
 Pass/Fail
 Spec
 Pass/Fail

 HbA1c %
 4.9 - 6.7
 Pass
 HbA1c %
 8.1 - 11.7
 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

8/29/2024

Date