



StatStrip Glucose Control Solution Level 1

Nova Ref Number 54520  
LifeScan Ref Number 02315402  
Lot Number 0424227241

Date of Manufacture 2024-08-14  
Exp Date 2026-08-14  
UDI# (01)00385480545204(11)240814(17)260814(10)0424227241

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57-65	60

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector  
Title

9/6/2024  
Date