OI-922 Page 55

Page 55 of 66 Rev. BL

Vialed and Packaged Test Strip Release

Test Strips

NOUAbiomedical CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:		NIPRO DIAGNOSTICS INC.
Description:		TEST STRIPS, NIPRO STATSTRIP TM GLUCOSE, PACKAGED
Catalog Number:	-	47940
Nova Biomedical P/N		47940
Target Range (mg/dL) SD (n	ng/dL) /CV (%)	
30-50	≤4 mg/dL	PASS FAIL
51-110	≤ 6.0%	PASS FAIL
111-150	≤ 5.0%	PASS FAIL
151-250	≤ 5.0%	PASS FAIL
251-400	≤ 5.0%	PASS FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method \geq 95%		PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%		PASS FAIL
Qty:		304
Lot No.		2124248309
Mfg. Date:		2024.09.04
Expiration Date:		2027.03.31

Quality Control Manager Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454

OI-922 Vialed and Packaged Test Strip Release Page 56 of 66 Rev. BL

Test Strips

<u>Al</u>

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2				
RELEASE TEST RESULTS				
ner:	NIPRO DIAGNOSTICS INC.			
4 :				
otion:	TEST STRIPS, NIPRO STATSTRIP TM GLUCOSE			
g Number:	47940 47940			
Biomedical P/N	47940			
•	2124248309			
(Spiked whole blood) Target Range:30-50				
SD (mg/dL) ≦4	YSI (reference, mg/dL))			
2.6	44.2			
(Spiked whole blood) Target Range:51-110				
CV% ≦6.0	YSI (reference, mg/dL))			
4.7	94.7			
(Spiked whole blood) Target Range:111-150 Mean (mg/dL) N=100 CV% ≤5.0 YSI (reference, mg/dL))				
CV% ≦5.0	YSI (reference, mg/dL))			
4.8	135.0			
(Spiked whole blood) Target Ra	nge:151-250			
CV% ≦5.0	YSI (reference, mg/dL))			
3.0	222.0			
(Spiked whole blood) Target Ra	nge:251-400			
CV% ≦5.0	YSI (reference, mg/dL))			
4.0	345.8			
	RELEASE TEST RESUMENTS Potion: Solumber: Solumber:			

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method \geq 95%	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%
97.47.	99.4%
- Marc Brown	e 9/11/24
Quality Control Manager	Date
Nova Biomedical Corporation	200 Prospect Street Waltham, MA 02454