



CERTIFICATE OF ANALYSIS

NCRT 4 Calibrator Pack
Lot Number 425328

Ref Number 15647
Exp Date 2025-11-07
Date of Manufacture 2024-08-07
UDI#(01)00385480156479(11)240807(17)251107(10)425328

| CAI A L/N 24225027 | | | Cal B L/N 24225028 | | |
|--------------------|-------------|------------------|--------------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| Na mmol/L | 129.1-130.9 | Pass | Na mmol/L | 59.4-60.6 | Pass |
| K mmol/L | 3.97-4.03 | Pass | K mmol/L | 9.8-10.2 | Pass |
| Cl mmol/L | 115.1-116.9 | Pass | Cl mmol/L | 60.4-61.6 | Pass |
| HCO3 mmol/L | 19.2-20.8 | Pass | HCO3 mmol/L | 9.6-10.4 | Pass |
| Li mmol/L | 0.49-0.51 | Pass | Li mmol/L | 4.95-5.05 | Pass |

| Cal H L/N 24220049 | | | Reference L/N 24220048 | | |
|---------------------|-------------|---------------|------------------------|-------------|------------------|
| | <u>Spec</u> | <u>Result</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| CO2 Releasing Agent | N/A | Present | KCl mol/L | 1.8-2.2 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

9/12/2024
Date