



CERTIFICATE OF ANALYSIS

NCRT 10 Calibrator Pack
Lot Number 424312

Ref Number 14608
Exp Date 2026-02-28
Date of Manufacture 2024-08-30
UDI# (01)00385480146081(11)240830(17)260228(10)424312

Cal A	L/N	24243012		Cal B	L/N	24243012	
		<u>Spec</u>	<u>Pass/Fail</u>			<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L		139.0 - 141.0	Pass	Na mmol/L		59.4 - 60.6	Pass
K mmol/L		3.97 - 4.03	Pass	K mmol/L		9.8 - 10.2	Pass
Cl mmol/L		113.6 - 115.4	Pass	Cl mmol/L		45.5 - 46.5	Pass
TCa mmol/L		1.94 - 2.06	Pass	TCa mmol/L		3.88 - 4.12	Pass
Li mmol/L		0.50 - 0.51	Pass	Li mmol/L		4.95 - 5.05	Pass

Ref L/N	24243012		
	<u>Spec</u>	<u>Result</u>	
Ca++ Releasing Agent	N/A	Present	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

9/12/2024
Date