



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	<u>PASS</u> / FAIL <u>PASS</u> / FAIL <u>PASS</u> / FAIL
Qty:	20,016
Lot No.	4924247129
Mfg. Date:	2024 - 09 - 03
Expiration Date:	2025 - 09 - 03
Unique Device Identifier (UDI):	(01)20385480433703 (11) 240903 (17) 250903 (10) 4924247129

Maureen Browne 9/16/2024
Quality Control Manager Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454