OI-922 Page 39 of 66

Rev. BL

Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:		LIFESCAN
Description:		TEST STRIPS, STATSTRIP <sup>™</sup> GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:		643577
Catalog Number:		42214
Nova Biomedical P/N		54515
One Touch P/N		02315302
Target Range (mg/dL) SD(mg/dL) /CV (%)		
30-50	≤4 mg/dL	PASS FAIL
51-110	≤ 6.0%	PASS FAIL
111-150	≤ 5.0%	PASS V FAIL
151-250	≤ 5.0%	PASS / FAIL
251-400	≤ 5.0%	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method ≥ 95%		PASS / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method $\geq$ 98%		PASS / FAIL
Qty:		36,288
Lot No.		0324239249
Date of Mfg. (DOM):		3074/08/30 (AAAA)
Expiration Date:		2016/08/26 (YYYY/MM/DD)
Unique Device Identifier (UDI):		(01)20385480422141
		(11) 2 4 0 8 2 6
		(17) 2 6 0 8 2 6
1 / 1		(10) 0324239249

Quality Control Manager

1/17/24 Date

Nova Biomedical Corporation

200 Prospect St

Waltham, MA 02454