



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and β -Ketone Control Solution Level 1
Lot Number 0421078301

Ref Number 46947
Exp Date 2023-09-19
Date of Manufacture 2021-03-19
UDI (01)00385480469470(11)210319(17)230919(10)0421078301

	Spec	Mean Result
Glucose mg/dL	57-65	64
Ketone mmol/L	0.1-0.2	0.2

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

4/13/2021
Date