



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and β -Ketone Control Solution Level 1
Ref Number 46947

Lot Number 0421153301
Exp Date 2023-12-02

Date of Manufacture 2021-06-02

UDI# (01)00385480469470(11)210602(17)231202(10)0421153301

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	63
Ketone mmol/L	0.1 - 0.2	0.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/14/2021
Date