



CERTIFICATE OF ANALYSIS

Bioprofile Flex Lactate Membrane
Lot Number 24256002

Part Number 42482
Exp Date 2025-09

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	27.64
Biopro Ctl L1	0.78 - 1.28	0.99
Biopro Ctl L3	4.25 - 5.75	5.31

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/19/2024

Date