



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and β -Ketone Control Solution Level 3

Lot Number 0421106303

Ref Number 46949

LifeScan Ref Number 02313001

Exp Date 2023-10-16

Date of Manufacture 2021-04-16

UDI# (01)00385480469494(11)210416(17)231016(10)0421106303

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275-325	319
Ketone mmol/L	2.5-3.5	3.0

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "Dardiff", written over a horizontal line.

Approval

QC Inspector

Title

5/11/2021

Date