



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and β -Ketone Control Solution Level 3
Nova Reference 46949

Lot Number 0421148303

Exp Date 2023-11-28

Date of Manufacture 2021-05-28

UDI# (01)00385480469494(11)210528(17)231128(10)0421148303

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	306
Ketone mmol/L	2.5 -3.5	3.4

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

6/29/2021

Date