



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A	
Description:	TEST STRIPS, STATSTRIP HEMOGLOBIN/HEMATOCRIT, PACKAGED 50 CT	
Catalog Number:	58786	
Nova Biomedical P/N	58786	
<u>Target Range (Hct)</u>	<u>CV (%)</u>	
28-36 %	≤ 8%	PASS / FAIL
38-46 %	≤ 8%	PASS / FAIL
48-56%	≤ 8%	PASS / FAIL
Qty:	20,034	
Lot No.	1524260249	
Date of Mfg. (DOM):	2024/09/16 (YYYY/MM/DD)	
Expiration Date:	2026/09/16 (YYYY/MM/DD)	
Unique Device Identifier (UDI):	(01)20385480587864 (11)240916 (17)260916 (10)1524260249	

Marc Bourne
 Quality Control Manager

9/23/24
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

