

CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose/Ketone Linearity Kit

Ref Number 46950

Lot Number 0521018990

Exp Date 2023-07-18

Date of Manufacture

2021-01-18

UDI# (01)00385480469500(11)210118(17)230718(10)0521018990

Level 1 L/N 052102530 Level 1 Exp Date 2023-07-25 Spec Glucose mg/dL 16-22	1 Mean Result 20	Level 2 L/N 0521032302 Level 2 Exp Date 2023-08-01 Spec Mean Result Glucose mg/dL 57-65 62 Ketone mmol/L 0.1-0.2 0.1
Level 3 L/N 0521035303 Level 3 Exp Date 2023-08-04 Spec Glucose mg/dL 105-125 Ketone mmol/L 0.7-1.0	Mean Result 118 0.9	Level 4 L/N 0521018304 Level 4 Exp Date 2023-07-18
Level 5 L/N 0521026305 Level 5 Exp Date 2023-07-26 Spec Glucose mg/dL 470-540 Ketone mmol/L 5.2-6.3	Mean Result 498 5.9	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. Analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector Title

....

4/1/2021

Date