nova biomedical

## CERTIFICATE OF ANALYSIS

Bio pHOx Cal Pack
Lot Number 24262027

Part Number 41874
Exp Date 2026-03

Std A L/N 24253055

| | Spec | Pass/Fail |
|---|---|---|
| Na mmol/L | 139.0 - 141.0 | Pass |
| HCO3 mmol/L | 19.6 - 20.4 | Pass |

Std B L/N 24254028

| | Spec | Pass/Fail |
|---|---|---|
| HCO3 mmol/L | 39.2 - 40.8 | Pass |

Std C L/N 24253052

| | Spec | Pass/Fail |
|---|---|---|
| pH @ 37C | 7.352 - 7.358 | Pass |

Std D L/N 24253059

| | Spec | Pass/Fail |
|---|---|---|
| pH @ 37C | 6.802 - 6.808 | Pass |
| Na mmol/L | 69.5 - 70.5 | Pass |

Reference L/N 24250051

| | Spec | Pass/Fail |
|---|---|---|
| KCl mol/L | 1.8 - 2.2 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

9/20/2024
Date