



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glucose Membrane
Lot Number 24261022

Part Number 42480
Exp Date 2025-12

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 90.00	45.52
Biopro Ctl L1	0.75 - 1.25	1.06
Biopro Ctl L3	12.75 - 17.25	15.35

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/27/2024

Date