



CERTIFICATE OF ANALYSIS

Nova StatStrip Lactate Control Solution Level 1
Lot Number 2721076301

Ref Number 47553
Exp Date 2023-09-17
Date of Manufacture 2021-03-17
UDI# (01)00385480475532(11)210317(17)230917(10)2721076301

| | <u>Spec</u> | <u>Mean Result</u> |
|----------------|-------------|--------------------|
| Lactate mmol/L | 0.5-0.7 | 0.7 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

5/3/2021
Date