



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Lactate Control Solution Level 1  
Lot Number 2722213301

Ref Number 47553  
Exp Date 2025-02-01  
Date of Manufacture 2022-08-01  
UDI# (01)00385480475532(11)220801(17)250201(10)2722213301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	0.5-0.7	0.7

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi M*

Approval

QC Inspector  
Title

9/20/2022  
Date