

CERTIFICATE OF ANALYSIS

Bioprofile Basic 2 Reagent Pack Lot Number 424941 Part Number 42237 Exp Date 2025-12-31

| Std A L/N | 24249041 | | Std B L/N | 24249041 | | Std C L/N | 24249041 | |
|---------------------|---------------------------------------|------------------------|---------------------|---------------------------------------|--------------------------|---------------------|---|------------------------|
| Gluc g/L Lac g/L | <u>Spec</u> 0.9 - 1.1 0.9 - 1.1 | Result Pass Pass | Gluc g/L Lac g/L | <u>Spec</u> 4.5 - 5.5 2.7 - 3.3 | Result Pass Pass | Gluc g/L Lac g/L | <u>Spec</u> 14.3 - 15.7 4.5 - 5.5 | Result Pass Pass |
| | Std D L/N | 24249041 | | | Flush L/N | 24249041 | | |
| Buf | fered Diluent | Spec N/A | Result Present | | Aqueous Salt Solution | Spec N/A | Result Present | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Gluc, Lac.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

Qc Inspector

Title

9/27/2024

Date