



CERTIFICATE OF ANALYSIS

Nova StatStrip Lactate Control Solution Level 2
Lot Number 2722214302

Ref Number 47554
Exp Date 2025-02-02
Date of Manufacture 2022-08-02
UDI# (01)00385480475549(11)220802(17)250202(10)2722214302

	Spec	Mean Result
Lactate mmol/L	5.9-6.7	6.2

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

9/20/2022

Date