



CERTIFICATE OF ANALYSIS

Nova 8 Ionized Magnesium Sensor
Lot Number 24256007

Ref Number 15265
Exp Date 2025-12-12

Date of Manufacture 2024-09-12

UDI# (01)10385480152652(11)240912(17)251212(10)24256007

| | <u>Spec*</u> | <u>Mean</u> |
|------------------------|--------------|-------------|
| Slope | 10.5 - 13.5 | 11.6 |
| Chem Control L1 mmol/L | 0.46 - 0.66 | 0.48 |
| Set A Level 2 mmol/L | 0.78 - 1.02 | 0.85 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

02/11/2025

Date