



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra/CCX Glucose Membrane  
Lot Number 22068016

Part Number 48552  
Exp Date 2023-06-09  
Date of Manufacture 2022-03-09  
UDI# (01)00385480485524(11)220309(17)230609(10)22068016

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	22.92
pHOx Ultra/CCX Ctl L4	73 - 87	78
pHOx Ultra/CCX Ctl L5	255 - 305	296

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

3/24/2022

Date