



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 24278037

Ref Number 53359
Exp Date 2026-03-23
Date of Manufacture 2024-09-23
UDI# (01)00385480533591(11)240923(17)260323(10)24278037

Cal A	L/N	24267050	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.353	N/A
pCO2	mmHg	25.0	N/A

Cal B	L/N	24267051	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
pO2	mmHg	167.0	N/A

Flush	L/N	24267052	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2	mmHg	49.0	N/A
pO2	mmHg	101.1	N/A

Reference	L/N	24267045	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/17/2024
Date