



CERTIFICATE OF ANALYSIS

STP pHox Plus Sodium Sensor
Lot Number 24274040

Ref Number 27810
Exp Date 2026-03-30
Date of Manufacture 2024-09-30
UDI# (01)00385480278102(11)240930(17)260330(10)24274040

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 - 10.50	Pass
Std C mV	-80.00 - +80.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/17/2024

Date