



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 700 Sample
Lot Number 24283065

Ref Number 65141

Exp Date 2026-04-01

Date of Manufacture 2024-10-01

UDI# (01)00385480651417(11)241001(17)260401(10)24283065

Table with 3 columns: Cal A L/N 24276030, Cal B L/N 24276031, Cal C L/N 24276032. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN mmol/L and mg/dL, and tHb g/dL with Spec, Pass/Fail, Assigned Value*, and Result columns.

Table with 3 columns: Cal D L/N 24276033, Flush L/N 24276034, Reference L/N 24275068. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, Assigned Value*, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature

Approval

QC Inspector

Title

10/22/2024

Date