



**CERTIFICATE OF ANALYSIS**

Prime ES Comp Plus Cal Cartridge 500 Sample  
Lot Number 24291047

Ref Number 62229  
Exp Date 2026-03-26  
Date of Manufacture 2024-09-26  
UDI# (01)00385480622295(11)240926(17)260326(10)24291047

Cal A	L/N	24270029		Cal B	L/N	24270028		Reference	L/N	24270027	
			<u>Spec</u>	<u>Pass/Fail</u>			<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L			134.1 - 135.9	Pass	Na mmol/L		76.5 - 77.5	Pass	KCl	1.8 - 2.2	Pass
K mmol/L			4.02 - 4.08	Pass	K mmol/L		9.80 - 10.20	Pass			
Cl mmol/L			114.1 - 115.9	Pass	Cl mmol/L		44.6 - 45.5	Pass			
iCa mmol/L			0.99 - 1.01	Pass	iCa mmol/L		1.98 - 2.02	Pass			
iMg mmol/L			0.41 - 0.51	Pass	iMg mmol/L		1.48 - 1.52	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/24/2024  
Date