

## **CERTIFICATE OF ANALYSIS**

STP pHOx/pHOx Basic/BiopHOx Air Detector Lot Number 24286007 Ref Number 21513 Date of Manufacture 2024-10-12

 Spec
 Pass/Fail

 AD1 mV
 1400 - 2600
 Pass

12 This 1400 2000 1 435

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

andang

Approval

QC Inspector Title

<u>10/25/2024</u> Date