



**CERTIFICATE OF ANALYSIS**

Nipro Glucose and  $\beta$ -Ketone Control Solution Level 2  
Lot Number 0424256302

Ref Number 11-116  
Exp Date 2027.03.31

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	105 - 125	111
Ketone mmol/L	0.7 - 1.1	0.9

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/24/2024

Date