

## **CERTIFICATE OF ANALYSIS**

## **Nova StatStrip Glucose Control Solution Level 2**

Nova Reference 54521 LifeScan Reference 02315502 Lot Number 0424274242 Date of Manufacture 2024-09-30 Exp Date 2026-09-30 UDI# (01)00385480545211(11)240930(17)260930(10)0424274242

Mean

<u>Spec</u>

Result

Glucose mg/dL

105 - 125

115

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/28/2024

Date