



CERTIFICATE OF ANALYSIS
Nova StatStrip Glucose Control Solution Level 1

Nova Reference 54520
LifeScan Reference 02315402
Lot Number 0424274241

Date of Manufacture 2024-09-30
Exp Date 2026-09-30
UDI# (01)00385480545204(11)240930(17)260930(10)0424274241

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awish M

Approval

QC Inspector

Title

10/28/2024

Date