



**CERTIFICATE OF ANALYSIS**  
**StatStrip Glucose Control Solution Level 3**

Nova Ref Number 54522  
LifeScan Ref Number 02315602  
Lot Number 0424275243

Date of Manufacture 2024-10-01  
Exp Date 2026-10-01  
(01)00385480545228(11)241001(17)261001(10)0424275243

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	302

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/28/2024  
Date