



CERTIFICATE OF ANALYSIS

Nova Vet Glucose and Ketone
Control Solution Mid Level
Lot Number 1030124222

Ref Number 52749
Exp Date 2027-02-09

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	99 - 114	108
Ketone mmol/L	0.8 - 1.0	0.9

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/28/2024
Date