



CERTIFICATE OF ANALYSIS

Nova Max Creatinine Control Solution Level 2
Lot Number 2740224276

Ref Number 63942
Exp Date 2026-10-02
Date of Manufacture 2024-10-02

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	2.90 - 3.90	3.39

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/30/2024
Date