



CERTIFICATE OF ANALYSIS

Glucose pHOX plus Membrane Kit
Lot Number 24298030

Part Number 28969
Exp Date 2026-01-24
Date of Manufacture 2024-10-24
UDI# (01)00385480289696(11)241024(17)260124(10)24298030

	<u>Spec*</u>	<u>Mean Result</u>
Slope	5.00 - 45.00	19.62
pHOX plus L3	76 - 90	84
Set A L1	≥ 470	514

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/31/2024
Date