



**CERTIFICATE OF ANALYSIS**

Allegro Lipids Test Cartridge  
Lot Number 24302010

Ref Number 54652  
Exp Date 2026-04-28  
Date of Manufacture 2024-10-28  
UDI# (01)00385480546522(11)241028(17)260428(10)24302010

| Ctl Level 1 |             |                  | Ctl Level 2 |             |                  |
|-------------|-------------|------------------|-------------|-------------|------------------|
|             | <u>Spec</u> | <u>Pass/Fail</u> |             | <u>Spec</u> | <u>Pass/Fail</u> |
| HDL mg/dL   | 28 - 50     | Pass             | HDL mg/dL   | 68 - 102    | Pass             |
| TCHOL mg/dL | 155 - 225   | Pass             | TCHOL mg/dL | 294 - 410   | Pass             |
| TGL mg/dL   | 60 - 136    | Pass             | TGL mg/dL   | 290 - 430   | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/6/2024  
Date