



CERTIFICATE OF ANALYSIS

StatSensor Creatinine Control Solution Level 2
Lot Number 5024276242

Ref Number 43922
Exp Date 2026-10-02
Date of Manufacture 2024-10-02
UDI# (01)00385480439220(11)241002(17)261002(10)5024276242

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	1.95 - 2.25	2.06

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

11/7/2024

Date