



**CERTIFICATE OF ANALYSIS**

STP pHox Plus Reference Sensor  
Lot Number 24307005

Ref Number 28954  
Exp Date 2026-05-02  
Date of Manufacture 2024-11-02  
UDI# (01)00385480289542(11)241102(17)260502(10)24307005

	<u>Spec</u>	<u>Pass/Fail</u>
Na Slope	9.00 - 11.00	Pass
ADT 4	≥ 220.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/8/2024  
Date