



CERTIFICATE OF ANALYSIS

StatSensor Creatinine Control Solution Level 1
Lot Number 5024276241

Ref Number 43921
Exp Date 2026-10-02
Date of Manufacture 2024-10-02
UDI# (01)00385480439213(11)241002(17)261002(10)5024276241

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	0.90 - 1.10	1.02

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector

Title

11/8/2024

Date