OI-922

Page 32 of 66

Rev. BL

Vialed and Packaged Test Strip Release

Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	
Customer.	NIA
	10 [1]
Description:	TEST STRIPS, NOVA MAX PRO
	CREATININE eGFR, 50 COUNT
Catalog Number:	63983
Nova Biomedical P/N	(2002
Nova Biomedical P/N	63983
Control Level 1: CV (%) \leq 15%	PASS FAIL
Control Level 2: CV (%) \leq 10%	PASS) FAIL
Qty:	
	540
Lot No.	2720021200
160 5	2730224309
Mfg. Date:	2024-11-04
Everinetic Det	
Expiration Date:	2025-11-04
II.	
Unique Device Identifier (UDI):	(01)20385480639839
	(11) 2 4 1 1 0 4
	(17) 2 5 1 1 0 4
	(10) \overline{a} $\overline{7}$ $\overline{3}$ $\overline{0}$ $\overline{2}$ \overline{a} $\overline{4}$ $\overline{3}$ $\overline{0}$ $\overline{9}$
Allen Range	

/////2029 Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454