



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

| | | |
|---|--|--------------------|
| Customer: | NIA | |
| Description: | TEST STRIPS, STATSTRIP CAPILLARY GLUCOSE ROW, PACKAGED 100 COUNT | |
| Catalog Number: | 61500 | |
| Nova Biomedical P/N | 61500 | |
| Target Range (mg/dL) | SD(mg/dL) /CV (%) | |
| 30-50 | ≤ 4 mg/dL | <u>PASS</u> / FAIL |
| 51-110 | ≤ 6.0% | <u>PASS</u> / FAIL |
| 111-150 | ≤ 5.0% | <u>PASS</u> / FAIL |
| 151-250 | ≤ 5.0% | <u>PASS</u> / FAIL |
| 251-400 | ≤ 5.0% | <u>PASS</u> / FAIL |
| Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95% | | <u>PASS</u> / FAIL |
| Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98% | | <u>PASS</u> / FAIL |
| Qty: | 3,706 | |
| Lot No. | 0324260309 | |
| Mfg. Date: | 2024-09-16 | |
| Expiration Date: | 2027-03-16 | |
| Unique Device Identifier (UDI): | (01)20385480615000 (11)240916 (17)270316 (10)0324260309 | |

Marc Broome
 Quality Control Manager

10/29/24
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454