



**CERTIFICATE OF ANALYSIS**

Acetate Conditioning Solution  
Lot Number 24312032

Part Number 34380  
Exp Date 2026/02

	Spec	Pass/Fail
pH	0.70-1.30	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "S. Smith".

Approval

QC Inspector  
Title

11/12/2024  
Date