



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamate Membrane Kit
Lot Number 24305024

Part Number 42484
Exp Date 2025-04

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	38.72
Biopro Ctl L4	0.27 - 0.77	0.45
Biopro Ctl L5	1.78 - 2.38	1.98

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

11/12/2024
Date