OI-922 Page 64 of 65 Rev. BG Vialed and Packaged Test Strip Release Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED
Catalog Number:	47486
Nova Biomedical P/N	47486
Target Range CV (%) SD (mmol/L)	
$0.8-2.5 \text{ mmol/L} \leq 6.5\% \pm 0.3$	PASS / FAIL
$3.0-5.0 \text{ mmol/L} \leq 5.5\% \qquad \pm 0.5$	PASS / FAIL
10-14 mmol/L $\leq 5.0\%$ ± 1.2	PASS / FAIL
Qty:	6,102
Lot No.	2622221249
Mfg. Date:	2022 - 08 - 09
Expiration Date:	2024 - 08 - 09
Unique Device Identifier (UDI):	(01)20385480 4 7 4 8 6 7
	$(11) \frac{2}{2} \frac{2}{4} \frac{08}{08} \frac{09}{08}$
	$ \begin{array}{c ccccccccccccccccccccccccccccccccccc$

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454