



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0424281301

Exp Date 2027-04-07

Date of Manufacture 2024-10-07

UDI # (01)00385480417419(11)241007(17)270407(10)0424281301

	Spec	Mean Result
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/12/2024  
Date