



CERTIFICATE OF ANALYSIS

Reagent Pack Autosampler Flex 2
Lot Number 24316032

Part Number 62450
Exp Date 2025-10-25
Date of Manufacturing 2024-10-25

Solution DP L/N		24299066	Solution S L/N		24299064
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
NaOCl %	0.43 - 0.73	Pass	pH	6.900 - 7.100	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO₃, pH, KCl, iCa, NH₄, K.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

11/14/2024

Date