



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ pH Sensor  
Lot Number 24306043

Part Number 48580  
Exp Date 2026-05-01

Date of Manufacture 2024-11-01

UDI# (01)00385480485807(11)241101(17)260501(10)24306043

|          | <u>Spec</u>   | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope    | 9.20 - 11.20  | Pass             |
| Std C mV | 0.00 ± 110.00 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

11/15/2024

Date