



**CERTIFICATE OF ANALYSIS**

Nova Creatinine / BUN / Hct Linearity STD Levels 1 - 4  
Lot Number 24298035

Ref Number 61656  
Exp Date 2025-10-22

Date of Manufacture 2024-10-22

UDI# (01)00385480616560(11)241022(17)251022(10)24298035

Level 1 L/N 24296026  
Level 1 Exp Date 2025-10-22

Level 2 L/N 24296037  
Level 2 Exp Date 2025-10-22

	<u>Spec</u>	<u>Pass/Fail</u>
BUN mg/dL	69.2 - 72.8	Pass
Creatinine mg/dL	9.5 - 9.9	Pass
Hct %	13 - 17	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
BUN mg/dL	46.8 - 49.2	Pass
Creatinine mg/dL	6.9 - 7.1	Pass
Hct %	28 - 32	Pass

Level 3 L/N 24297023  
Level 3 Exp Date 2025-10-23

Level 4 L/N 24297024  
Level 4 Exp Date 2025-10-23

	<u>Spec</u>	<u>Pass/Fail</u>
BUN mg/dL	23.4 - 24.6	Pass
Creatinine mg/dL	4.5 - 4.7	Pass
Hct %	44 - 48	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
BUN mg/dL	4.9 - 5.1	Pass
Creatinine mg/dL	0.98 - 1.02	Pass
Hct %	62 - 66	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/13/2024  
Date