

CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge Lot Number 24318013

rridge Ref Number 54215 Exp Date 2025-11-13 Date of Manufacture 2024-11-13 UDI# (01)00385480542159(11)241113(17)251113(10)24318013

 Ctl Level 1
 Ctl Level 2

 Spec
 Pass/Fail
 Spec
 Pass/Fail

 HbA1c %
 4.7 - 6.9
 Pass
 HbA1c %
 7.7 - 12.1
 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector Title

11/18/2024 Date